



# TTESTATION OF CONFORM

Certificate No: MD-370

## ICE GRUP PLASTIK ANONIM ŞİRKETİ

Barbaros Mahallesi Mor Sümbül Sokak No:1 Varyap Meridian Business I Blok Kat:8 Ofis:110 Ataşehir / İSTANBUL

Manufacturing Site:

Çerkeşli Osb Mah. 6. İmes 10.Cad No:10 Çerkeşli Köyü Dilovası Kocaeli

product, identified below, found to meet the applicable requirements of Regulation (EU) 2017/745 for The manufacturer's technical documentation (Dated 27.08.2021), manufacturing facilities and the Class I devices, including the General Safety and Performance Requirements in Annex I of the regulation based on assessment results and evaluation of relevant test reports.

Identification of the Product

Brand Name: ICE PRO MASK, Model: MB3R

**UDI-DI Number:** 8683465839MB3RHU

Classification: Type IIR

loops and nose bridge. The mask is available in 1 size and have black, blue and white colour variants Medical masks, as a medical device, manufactured from spunbound and meltblown fabrics, with ear

For more details, refer to the evaluation report provided to the manufacturer, dated 29.11.2021 and number MD-TR-370.

# The following harmonised standards have been applied:

EN 14683+AC:2019, Medical face masks - Requirements and test methods.

manufacturer according to Class I product requirements defined in the medical device regulation EU 2017/745 to fix the CE mark on the product, identified above. The manufacturer shall also issue an appropriate EU Declaration of Conformity according to Medical Devices regulation (EU) 2017/745 This certification is based on the voluntary scheme for verification of obligations fulfilled by the

This certificate is initially issued on 29/11/2021, will be valid for 1 year from the issue date, without any change in the design and manufacturing process of the product.

ISTANBUL - 29.11.2021

Verify the validity with the QR Code

UNIVERSAL CERTIFICATION Suat KACMAZ Director

This certificate will be in the absence of any changes in standard and legal terms, and with the surveillance annually following the surveillance audits, updating the publicationdate without changing the certificate number. audits to be concuted



## TECHNICAL EVALUATION REPORT

REPORT DATE / NO: 29.11.2021 / MD-TR-370

Manufacturer: ICE GRUP PLASTİK ANONİM ŞİRKETİ

Address: Barbaros Mahallesi Mor Sümbül Sokak No:1 Varyap Meridian Business I Blok Kat:8

Ofis:110 Ataşehir / ISTANBUL

Manufacturing Site: Çerkeşli Osb Mah. 6. İmes 10.Cad No:10 Çerkeşli Köyü Dilovası Kocaeli

SRN: 8683465

2017/745, Medical Device Regulation for Class I products upon the manufacturer request on a voluntary base. ZA of harmonised standard EN 14683+AC:2019 and the General Safety and Performance Requirements of EU The medical masks (non steril) manufactured by the above manufacturer, are evaluated based on the Annex

fabrics, with ear loops and nose bridge. The mask is available in 1 size and have black, blue and white colour Product Description: Medical masks, as a medical device, manufactured from spunbound and meltblown

Trademark: ICE PRO MASK Model: MB3R

UDI - DI: 8683465839MB3RHU

Sample Photos







the EN 14683+AC:2019 standard. evaluated and the samples taken from the manufacturing site during the assessment visit conducted by Universal Certification on 03/08/2021 are tested according to the test requirements referred in Annex ZA of As a third party evaluation, the technical file and other official documentation provided by the manufacturer is

See EN ISO/IEC 17025 Accredited Laboratory Test Reports,

number 21032061-Ing. Annex I: Test report by Ekoteks Laboratuvar ve Gözetim Hizmetleri A.Ş. dated 01.11.2021 with report

Engineering and Biotechnology Institute dated 19.11.2021 with report number 16563500-125.05-150/7888. Annex II: Test report by TUBITAK (The Scientific and Technological Research Council of Turkey) Genetic

reponsibility of the manufacturer covered under EU 2017/745 Medical Device Regulation. The manufacturer shall continue to fulfil all responsibilities for Class I products under EU 2017/745 Medical Device Regulation. This report or the issued certificate, in case the report is positive, does not take over or change the sole





The results of the evaluation are as follows:

# A- Review of the technical file, Marking and Information Requirements

In the evaluation of the technical and other official documentation the following information was

number 8683465839MB3L9. system with the SRN number 8683465. The product is registered in the UDI system with UDI-DI documentation on the mentioned product. The applicant is a registered manufacturer on the UDI The manufacturer follows the requirements of the EU 2017/745 to establish its technical

system established is certified based on ISO 13485 standard. identification of applicable general safety and performance requirements, strategy for regulatory compliance, product realisation and management of modifications in the device. The management The manufacturer established a quality management system including a risk management,

conditions, warnings, limitations, controls before use, the intended use of the product, disposal information for use is delivered to the users as printed on the cardboard box. The manufacturer, model of the product, UDI-DI identification, the lot number of the product, shelf life, storage method are given within this information. Section 23 of the EU 2017/745 regulation and also based on EN 14683+AC:2019 Clause 6. The The marking and the information supplied for the consumers are evaluated based on the Annex I

En\*\*\* according to Article 15 of the regulation. The manufacturer has employed a person in charge responsible for regulatory compliance Me\*\*

to quality management systems, risk management, post-market surveillance according to the Article 8 of MDR, the manufacturer claims the fulfilment of requirements relating investigations, clinical evaluation or post-market clinical follow-up ('PMCF'). The manufacturer is using the harmonised standard EN 14683+AC:2019 for the product and systems, clinical

#### **B-** Product Test Results

and the results are evaluated below. More information can be gathered on tests from the referred The tests referenced in Annex ZA are conducted on the samples taken from the manufacturing site

### 1. Biocompatibility (Annex II)

irritation according to EN 10993-10:2010 standard and found to have no irritation affect. addition to the manufacturer's declaration on biocompatibility, the samples are tested for declares that the produced masks are complies with the biocompatibility requirements and their biocompatibility. The manufacturer claims that the request and evaluation of proofs for biocompatibility of the goods is an essential part of the procurement policy and have authorised responsible staff members for ensuring the success of this policy. established a mechanism for the evaluation of raw materials or semi-finished goods on In the evaluation of the technical file, it was observed that the manufacturer has

### 2. Bacteria Filtration Efficiency

annexed test report. the results of the incubation of samples taken in different particule sizes are shown in the minutes with a test setup defined in the Annex B of EN 14683+AC:2019 standard. With At least 5 samples are subjected to a bacteria aerosol with a flow rate of 28.3 L/min for 2

classes are shown below; The minimum bacteria filtration efficiency performance required by each performance



UFR-896 01.06.2021 Rev.00

Page 2|4



Test	Type I*	Type II	Type IIR
<b>Bacterial Filtration</b>	2		
Efficiency (BFE), (%)	1/95	≥ 98	≥ 98

the risk of spread of infections particularly in epidemic or pandemic situations. Type I other medical settings with similar requirements. masks are not intended for use by healthcare professionals in an operating room or in \* Type I medical face masks should only be used for patients and other persons to reduce

efficiency performance of the masks is classifified as Type IIR filtration efficiency is given as 98,6%. According to this result, the bacteria filtration According to the evaluation of the results of 5 samples tested, the minimum bacteria

reported as a confidence parameter of the test result are meaningful. It was observed that the avarage positive control values and negative control value is also

## 3. Microbial Cleanliness (Bioburden)

classes (Type I, Type II and Type IIR). reported as 2 For this test result the samples complies the requirement for all performance In the evaluation of the test result, the maximum count of the colony forming unit is all performance class of masks according to the test result based on ISO 11737-1 standard. It is expected to have the number of colony forming units per gram to be lower than 30 for

#### 4. Differentail Pressure

for Type IIR class not to be higher than 60 Pa/cm<sup>2</sup> the expected result for Type I and Type II classes is not to be higher than 40 Pa/cm<sup>2</sup> and The test is conducted to measure the breathing resistance as the differential pressure and

the samples complies the requirement for all performance classes (Type I, Type II and According to the test results, the highest differential pressure measured is 31,4 Pa/cm<sup>2</sup> and

#### 5. Splash Resistance Pressure

to be equal or higher than 16kpa for the Type 2R class. In the test, done according to ISO 22609:2004 the product's splash resistance is expected

All 32 samples tested were able to provide Type IIR performances as 16kPa resistance.



UFR-896 01.06.2021 Rev.00



#### C- Summary and Conclusion

Evaluation	Requirement	Result	Classification
Biocompatibility Irritation	Not Irritant	Not Irritant	,
Bacterial Filtration Efficiency (BFE), (%)	≥ 95 % – Type I ≥ 98 % – Type II ≥ 98 % – Type IIR	98,6 %	Type I Type II
Differential pressure (Pa/cm2)	< 40 – Type I < 40 – Type II < 60 – Type IIR	31,4	Type I Type II Type IIR
Splash resistance pressure (kPa)	Not Required – Type I  Not Required – Type II  ≥ 16 – Type IIR	≥16	Type IIR
Microbial cleanliness (cfu/g)	≤30 – Type I ≤30 – Type II ≤30 – Type IIR	2	Type I Type II Type IIR
Overall Performance Classification	sification		Type IIR

#### - End of Report -



UFR-896 01.06.2021 Rev.00